
Anode response comparison H8711MOD
tsukuba (16ch scan) vs. hamamatsu
Tsukuba, pixel centers
Hamamatsu
0
0.5
1
1.5
2
Corner
Middle
Left/Right
Top/Bottom


Anode response comparison H8711MOD
tsukuba (16ch scan) vs. hamamatsu
Tsukuba, pixel centers
Hamamatsu
Region:
Corner
χ²/ndf 1248E+05/ 31
P1 0.8637 ± 0.4159E-03
Region:
Middle
χ²/ndf 9504. / 31
P1 1.222 ± 0.6344E-03
Region:
Left/Right
χ²/ndf 1027E+05/ 31
P1 1.039 ± 0.5064E-03
Region:
Top/Bottom
χ²/ndf 5215. / 31
P1 0.9470 ± 0.4657E-03